



Standard Guide for Making Quality Nondestructive Assay Measurements¹

This standard is issued under the fixed designation C1592/C1592M; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reappraisal.

1. Scope

1.1 This guide is a compendium of Quality Measurement Practices for performing measurements of radioactive material using nondestructive assay (NDA) instruments. The primary purpose of the guide is to assist users in arriving at quality NDA results, that is, results that satisfy the end user's needs. This is accomplished by providing an acceptable and uniform basis for the collection, analysis, comparison, and application of data. The recommendations are not compulsory or prerequisites to achieving quality NDA measurements, but are considered contributory in most areas.

1.2 This guide applies to the use of NDA instrumentation for the measurement of nuclear materials by the observation of spontaneous or stimulated nuclear radiations, including photons, neutrons, or the flow of heat. Recommended calibration, operating, and assurance methods represent guiding principles based on current NDA technology. The diversity of industry-wide nuclear materials measurement applications and instrumentation precludes discussion of specific measurement situations. As a result, compliance with practices recommended in this guide must be based on a thorough understanding of contributing variables and performance requirements of the specific measurement application.

1.3 Selection of the best instrument for a given measurement application and advice on the use of this instrument must be provided by a qualified NDA professional following guidance provided in Guide C1490. This guide is to be used as a reference, and to supplement the critical thinking, professional skill, expert judgment, and experimental test and verification needed to ensure that the instrumentation and methods have been properly implemented.

1.4 The intended audience for this guide includes but is not limited to Management, Auditor Support, NDA Qualified Instrument Operators, NDA Technical Specialists, and NDA Professionals.

1.5 The values stated in either SI units or inch-pound units are to be regarded separately as standard. The values stated in each system may not be exact equivalents; therefore, each system shall be used independently of the other. Combining values from the two systems may result in non-conformance with the standard.

2. Referenced Documents

2.1 *ASTM Standards:*²

- C1030 Test Method for Determination of Plutonium Isotopic Composition by Gamma-Ray Spectrometry
- C1133 Test Method for Nondestructive Assay of Special Nuclear Material in Low-Density Scrap and Waste by Segmented Passive Gamma-Ray Scanning
- C1207 Test Method for Nondestructive Assay of Plutonium in Scrap and Waste by Passive Neutron Coincidence Counting
- C1215 Guide for Preparing and Interpreting Precision and Bias Statements in Test Method Standards Used in the Nuclear Industry
- C1221 Test Method for Nondestructive Analysis of Special Nuclear Materials in Homogeneous Solutions by Gamma-Ray Spectrometry
- C1254 Test Method for Determination of Uranium in Mineral Acids by X-Ray Fluorescence
- C1268 Test Method for Quantitative Determination of ²⁴¹Am in Plutonium by Gamma-Ray Spectrometry
- C1316 Test Method for Nondestructive Assay of Nuclear Material in Scrap and Waste by Passive-Active Neutron Counting Using ²⁵²Cf Shuffler
- C1455 Test Method for Nondestructive Assay of Special Nuclear Material Holdup Using Gamma-Ray Spectroscopic Methods
- C1458 Test Method for Nondestructive Assay of Plutonium, Tritium and ²⁴¹Am by Calorimetric Assay
- C1490 Guide for the Selection, Training and Qualification of Nondestructive Assay (NDA) Personnel
- C1493 Test Method for Non-Destructive Assay of Nuclear Material in Waste by Passive and Active Neutron Counting Using a Differential Die-Away System

¹ This guide is under the jurisdiction of ASTM Committee C26 on Nuclear Fuel Cycle and is the direct responsibility of Subcommittee C26.10 on Non Destructive Assay.

Current edition approved June 1, 2009. Published July 2009. Originally approved in 2004. Last previous edition approved in 2004 as C1592 – 04. DOI: 10.1520/C1592_C1592M-09.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.