This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.

Designation: D857 – 17

Standard Test Method for Aluminum in Water[1]

This standard is issued under the fixed designation D857; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope*

1.1 This test method covers the direct flame atomic absorption determination of aluminum in the nitrous oxide-acetylene flame.

1.2 This test method is applicable to waters containing dissolved and total recoverable aluminum in the range from 0.5 to 5.0 mg/L. Aluminum concentrations as high as approximately 50 mg/L can be determined using this test method without dilution. However, no precision and bias data are available for concentrations greater than 5.0 mg/L.

1.3 This test method was tested on reagent, natural, and potable waters. It is the user's responsibility to ensure the validity of this test method for waters of untested matrices.

1.4 The same digestion procedure may be used to determine total recoverable nickel (Test Methods D1886), chromium (Test Methods D1687), cobalt (Test Methods D3558), copper (Test Methods D1688), iron (Test Methods D1068), lead (Test Methods D3559), manganese (Test Methods D858), and zinc (Test Methods D1691).

1.5 Precision and bias data have been obtained on reagent, natural, and potable waters. It is the responsibility of the user to ensure the validity of this test method on untested matrices.

1.6 The values stated in SI units are to be regarded as standard. The values given in parentheses are mathematical conversion to inch-pound units that are provided for information only and are not considered standard.

1.7 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific hazard statements, see 7.5, 8.4, and Note 2.

1.8 Former Test Methods A (Fluorometric) and B and C (Spectrophotometric) were discontinued. Refer to Appendix X1 for historical information.

1.9 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 *ASTM Standards:*[2]

D858 Test Methods for Manganese in Water
D1066 Practice for Sampling Steam
D1068 Test Methods for Iron in Water
D1129 Terminology Relating to Water
D1193 Specification for Reagent Water
D1687 Test Methods for Chromium in Water
D1688 Test Methods for Copper in Water
D1691 Test Methods for Zinc in Water
D1886 Test Methods for Nickel in Water
D2777 Practice for Determination of Precision and Bias of Applicable Test Methods of Committee D19 on Water
D3370 Practices for Sampling Water from Closed Conduits
D3558 Test Methods for Cobalt in Water
D3559 Test Methods for Lead in Water
D4841 Practice for Estimation of Holding Time for Water Samples Containing Organic and Inorganic Constituents
D5673 Test Method for Elements in Water by Inductively Coupled Plasma—Mass Spectrometry
D5810 Guide for Spiking into Aqueous Samples
D5847 Practice for Writing Quality Control Specifications for Standard Test Methods for Water Analysis

3. Terminology

3.1 *Definitions:*

3.1.1 For definitions of terms used in this standard, refer to Terminology D1129.

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *continuing calibration blank, n*—a solution containing no analytes (of interest) which is used to verify blank response and freedom from carryover.

[1] These test methods are under the jurisdiction of ASTM Committee D19 on Water and are the direct responsibility of Subcommittee D19.05 on Inorganic Constituents in Water.

Current edition approved June 1, 2017. Published June 2017. Originally approved in 1945. Last previous edition approved in 2012 as D857 – 12. DOI: 10.1520/D0857-17.

[2] For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

*A Summary of Changes section appears at the end of this standard

Copyright © ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959. United States