



Standard Test Method for Separation and Collection of Particulate and Water-Soluble Gaseous Fluorides in the Atmosphere (Filter and Impinger Method)¹

This standard is issued under the fixed designation D3267; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method provides a procedure for separation and collection on a chemically treated filter and in an impinger of particulate and water-soluble gaseous fluoride in the atmosphere. The sampling rate may vary from 30 L/min (1.0 ft³/min) to 15 L/min (0.5 ft³/min) for longer sampling periods depending on the atmospheric fluoride concentration. This test method is not intended to be applied to gaseous fluorine compounds that are not water-soluble

1.2 There are several limitations of the test method:

1.2.1 Although the acid-treated, medium retentive, prefilter has been shown to allow passage of HF, it will restrict passage of particulate matter only as small as about 1 μm. Thus, smaller particulate matter may pass through the filter and be collected in or pass through the impingers.

1.2.2 The maximum sample volume to be taken using this test method prior to changing the acid-treated prefilter is recommended as 12 m³. This recommendation is made to minimize any effects due to particulate matter build-up.

1.3 The values stated in SI units are to be regarded as the standard. The inch-pound units given in parentheses are for information only.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific precautionary statements, see 7.2 and 7.4.6.

2. Referenced Documents

2.1 *ASTM Standards:*²

D1071 Test Methods for Volumetric Measurement of Gaseous Fuel Samples

D1193 Specification for Reagent Water

D1356 Terminology Relating to Sampling and Analysis of Atmospheres

D1357 Practice for Planning the Sampling of the Ambient Atmosphere

D2009 Practice for Collection by Filtration and Determination of Mass, Number, and Optical Sizing of Atmospheric Particulates (Withdrawn 1995)³

D3268 Test Method for Separation and Collection of Particulate and Gaseous Fluorides in the Atmosphere (Sodium Bicarbonate-Coated Glass Tube and Particulate Filter Method)

D3269 Test Methods for Analysis for Fluoride Content of the Atmosphere and Plant Tissues (Manual Procedures) (Withdrawn 2010)³

D3270 Test Methods for Analysis for Fluoride Content of the Atmosphere and Plant Tissues (Semiautomated Method)

E337 Test Method for Measuring Humidity with a Psychrometer (the Measurement of Wet- and Dry-Bulb Temperatures)

3. Terminology

3.1 *Definitions*—For definitions of terms used in this test method, refer to Terminology D1356.

4. Summary of Test Method

4.1 Air is drawn through an air inlet tube (see Practice D1357) and is first passed through an acid-treated prefilter to remove particulate matter that may contain fluoride and then through an impinger to remove water-soluble fluorides.

¹ This test method is under the jurisdiction of ASTM Committee D22 on Air Quality and is the direct responsibility of Subcommittee D22.03 on Ambient Atmospheres and Source Emissions.

Current edition approved April 1, 2012. Published May 2012. Originally approved in 1973. Last previous edition approved in 2005 as D3267 – 91 (2005). DOI: 10.1520/D3267-12.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ The last approved version of this historical standard is referenced on www.astm.org.