



Standard Test Method for Benzene in Hydrocarbon Solvents by Gas Chromatography¹

This standard is issued under the fixed designation D4367; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope*

1.1 This test method covers the determination by gas chromatography of benzene at levels from 0.01 to 1 volume % in hydrocarbon solvents.

NOTE 1—For benzene levels lower than 0.01 volume %, use Test Method [D6229](#).

1.2 The values stated in SI units are to be regarded as the standard. The values given in parentheses are for information only.

1.3 For purposes of determining conformance of an observed or a calculated value using this test method to relevant specifications, test result(s) shall be rounded off “to the nearest unit” in the last right-hand digit used in expressing the specification limit, in accordance with the rounding-off method of Practice [E29](#).

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

1.5 For hazard information and guidance, see the supplier’s Material Safety Data Sheet. For specific hazard statements, see Section [7](#).

2. Referenced Documents

2.1 ASTM Standards:²

[D6229](#) Test Method for Trace Benzene in Hydrocarbon Solvents by Capillary Gas Chromatography

[E29](#) Practice for Using Significant Digits in Test Data to Determine Conformance with Specifications

[E300](#) Practice for Sampling Industrial Chemicals

¹ This test method is under the jurisdiction of ASTM Committee [D01](#) on Paint and Related Coatings, Materials, and Applications and is the direct responsibility of Subcommittee [D01.35](#) on Solvents, Plasticizers, and Chemical Intermediates.

Current edition approved July 1, 2012. Published September 2012. Originally approved in 1984. Last previous edition approved in 2007 as D4367 – 02 (2007). DOI: 10.1520/D4367-02R12.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard’s Document Summary page on the ASTM website.

3. Summary of Test Method

3.1 An internal standard, methyl ethyl ketone (MEK), is added to the material and then introduced into a gas chromatograph equipped with two columns connected in series. The specimen passes first through a column packed with the nonpolar phase, methyl silicone, which separates the components by boiling point. After octane has eluted, the flow through the nonpolar column is reversed, flushing out the components heavier than octane. The octane and lighter components then pass through a column with the highly polar phase, 1,2,3-tris(2-cyanoethoxy)propane, that separates the aromatic and nonaromatic compounds. The eluted components are detected by a conventional detector and recorded on a strip chart. The peak areas are measured and the concentration of each component is calculated by reference to the internal standard.

4. Significance and Use

4.1 Benzene is classed as a toxic and carcinogenic material. A knowledge of the concentration of this compound may be an aid in evaluating the possible health hazards to persons handling and using hydrocarbon solvents, but this test method is not intended to evaluate such hazards.

5. Apparatus

5.1 *Chromatograph*—Any gas chromatographic instrument that has a backflush system and flame ionization detector and that can be operated at the conditions given in [Table 1](#). The detector-recorder combination must produce a 4-mm deflection for a 1- μ L specimen containing 0.05 volume % MEK when operated at maximum sensitivity.

5.2 *Columns*, one 0.8-m (2.5-ft) length of 3.2-mm ($\frac{1}{8}$ -in.) outside diameter stainless steel tubing and one 4.6-m (15-ft) length of 3.2-mm ($\frac{1}{8}$ -in.) outside diameter stainless steel tubing.

5.3 *Recorder, Strip Chart*—Potentiometer with a full-scale deflection of 1 mV, a full-scale response time of 2 s or less, and a maximum noise level of ± 0.3 % of full scale.

5.4 *Microsyringe*, 5- μ L capacity.

5.5 *Pipets*, measuring 1 and 2 mL, graduated in 0.01 mL; 5, 10, and 20-mL capacity.

5.6 *Flasks*, volumetric, 25 and 100-mL capacity.

*A Summary of Changes section appears at the end of this standard