



Standard Practice for Calibration of Ozone Monitors and Certification of Ozone Transfer Standards Using Ultraviolet Photometry¹

This standard is issued under the fixed designation D5110; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice covers a means for calibrating ambient, workplace, or indoor ozone monitors, and for certifying transfer standards to be used for that purpose.

1.2 This practice describes means by which dynamic streams of ozone in air can be designated as primary ozone standards.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* See Section 8 for specific precautionary statements.

2. Referenced Documents

2.1 *ASTM Standards:*²

D1356 Terminology Relating to Sampling and Analysis of Atmospheres

D3195 Practice for Rotameter Calibration

D3249 Practice for General Ambient Air Analyzer Procedures

D3631 Test Methods for Measuring Surface Atmospheric Pressure

D5011 Practices for Calibration of Ozone Monitors Using Transfer Standards

E220 Test Method for Calibration of Thermocouples By Comparison Techniques

E591 Practice for Safety and Health Requirements Relating

to Occupational Exposure to Ozone (Withdrawn 1990)³
E644 Test Methods for Testing Industrial Resistance Thermometers

3. Terminology

3.1 *Definitions:* For definitions of terms used in this practice, refer to Terminology **D1356**.

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *primary standard*—a standard directly defined and established by some authority, against which all secondary standards are compared.

3.2.2 *secondary standard*—a standard used as a means of comparison, but checked against a primary standard.

3.2.3 *standard*—an accepted reference sample or device used for establishing measurement of a physical quantity.

3.2.4 *transfer standard*—a type of secondary standard. It is a transportable device or apparatus that, together with operational procedures, is capable of reproducing pollutant concentration or producing acceptable assays of pollutant concentrations.

3.2.5 *zero air*—purified air that does not contain ozone, and does not contain any other component that may interfere with the measurement (see 7.1).

4. Summary of Practice

4.1 This practice is based on the photometric assay of ozone (O_3) concentrations in a dynamic flow system. The concentration of O_3 in an absorption cell is determined from a measurement of the amount of 253.7 nm light absorbed by the sample. This determination requires knowledge of (1)⁴ the absorption coefficient of O_3 at 253.7 nm, (2) the optical path length through the sample, (3) the transmittance of the sample at a wavelength of 253.7 nm, and (4) the temperature and pressure of the sample. The transmittance is defined as the ratio:

$$I/I_o$$

¹ This practice is under the jurisdiction of ASTM Committee D22 on Air Quality and is the direct responsibility of Subcommittee D22.03 on Ambient Atmospheres and Source Emissions.

Current edition approved Oct. 1, 2010. Published November 2010. Originally approved in 1990. Last previous edition approved in 2004 as D5110 - 98(2004). DOI: 10.1520/D5110-98R10.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ The last approved version of this historical standard is referenced on www.astm.org.

⁴ The boldface numbers in parentheses refer to the references listed at the end of this practice.