



Designation: D 5771 – 02

An American National Standard



Designation: 444/99

Standard Test Method for Cloud Point of Petroleum Products (Optical Detection Stepped Cooling Method)¹

This standard is issued under the fixed designation D 5771; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

INTRODUCTION

This test method describes an alternative procedure for the determination of cloud point of petroleum products Test Method D 2500 using an automatic apparatus. The temperature results from this test method have been found to be equivalent to Test Method D 2500. When specification requires Test Method D 2500, do not substitute this test method or any other method without obtaining comparative data and agreement from the specifier.

1. Scope

1.1 This test method describes the determination of the cloud point of petroleum products and biodiesel fuels that are transparent in layers 40 mm in thickness, by an automatic instrument using an optical device.

1.2 This test method covers the range of temperatures from –40 to +49°C with temperature resolution of 0.1°C.

1.3 The values stated in SI units are to be regarded as the standard. The values given in parentheses are for information only.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:

- D 2500 Test Method for Cloud Point of Petroleum Products²
- D 4057 Practice for Manual Sampling of Petroleum and Petroleum Products³
- D 4177 Practice for Automatic Sampling of Petroleum and Petroleum Products³

2.2 IP Standard:⁴

IP219 Test Method for Cloud Point of Petroleum Products⁴

2.3 Other Standard:⁵

IEC 751 Industrial Platinum Resistance Thermometer Sensors

3. Terminology

3.1 Definition:

3.1.1 *cloud point, n—in petroleum products and biodiesel fuels*, the temperature of a liquid specimen when the smallest observable cluster of wax crystals first appears upon cooling under prescribed conditions.

3.1.1.1 *Discussion*—The cloud point appears when the temperature of the specimen is low enough to cause wax crystals to precipitate. In homogeneous liquids, the cloud is always noted first at the location in the specimen where the specimen temperature is the lowest. This is typically at the lower portion of the test jar when using the apparatus described in Test Method D 2500.

3.1.2 *biodiesel, n—fuel comprised of mono-alkyl esters of long chain fatty acids derived from vegetable oils or animal fats, designated B 100.*

3.1.2.1 *Discussion*—Biodiesel is typically produced by a reaction of vegetable oil or animal fat with an alcohol such as methanol or ethanol in the presence of a catalyst to yield mono-esters and glycerin. The fuel typically may contain up to 14 different types of fatty acids that are chemically transformed into fatty acid methyl esters (FAME).

¹ This test method is under the jurisdiction of ASTM Committee D02 on Petroleum Products and Lubricants and is the direct responsibility of Subcommittee D02.07 on Flow Properties.

Current edition approved April 10, 2002. Published July 2002. Originally published as D 5771–95. Last previous edition D 5771–95.

² *Annual Book of ASTM Standards*, Vol 05.01.

³ *Annual Book of ASTM Standards*, Vol 05.06.

⁴ Available from the Institute of Petroleum, 61 New Cavendish St., London, England W1M 8AR.

⁵ Available from International Electrotechnical Commission, P.O. 131, 3 Rue de Varembe, CH-Geneve 1211–2.0.