



Standard Test Method for Determination of Gaseous and Particulate Polycyclic Aromatic Hydrocarbons in Ambient Air (Collection on Sorbent-Backed Filters with Gas Chromatographic/Mass Spectrometric Analysis)¹

This standard is issued under the fixed designation D 6209; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (€) indicates an editorial change since the last revision or reapproval.

€¹ NOTE—Paragraph 11.2.2 was corrected editorially in December 1998.

1. Scope

1.1 This test method² specifies sampling, cleanup, and analysis procedures for the determination of polycyclic aromatic hydrocarbons (PAH) in ambient air.

1.2 This test method is designed to collect both gas-phase and particulate-phase PAH and to determine them collectively.

1.3 This test method is a high-volume sampling (100 to 250 L/min) method capable of detecting PAH at sub-nanograms per cubic meter (ng/m³) concentrations with sampling volumes up to 350 m³ of air.

1.4 This test method has been validated for sampling periods up to 24 h.

1.5 Precision and bias under normal conditions can be expected to be ±35 to 50 %.

1.6 This test method describes a sampling and analysis procedure for PAH that involves collection from air on a combination fine-particle filter and sorbent trap and subsequent analysis by gas chromatography/mass spectrometry (GC/MS).

1.7 The range of this test method is approximately 0.05 to 1000 ng/m³ of air sampled.

1.8 The values stated in SI units shall be regarded as standard.

1.9 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applica-*

bility of regulatory limitations prior to use. See also Section 8 for additional safety precautions.

2. Referenced Documents

2.1 ASTM Standards:

D 1356 Terminology Relating to Sampling and Analysis of Atmospheres³

D 1357 Practice for Planning the Sampling of the Ambient Atmosphere³

D 3631 Test Methods for Measuring Surface Atmospheric Pressure³

E 1 Specification for ASTM Thermometers⁴

3. Terminology

3.1 *Definitions*— For definitions of terms used in this test method, refer to Terminology D 1356.

3.2 Definitions of Terms Specific to This Standard:

3.2.1 *sampling efficiency (SE), n*—ability of the sampler to trap and retain PAH. The percent SE is the percentage of the analyte of interest collected and retained by the sampling medium when it is introduced into the air sampler and the sampler is operated under normal conditions for a period of time equal to or greater than that required for the intended use.

3.2.2 *dynamic retention efficiency, n*—ability of the sampling medium to retain a given PAH that has been added to the sorbent trap in a spiking solution when air is drawn through the sampler under normal conditions for a period of time equal to or greater than that required for the intended use.

4. Summary of Test Method

4.1 Sampling:

4.1.1 An air sample is collected directly from the ambient atmosphere by pulling air at approximately 225 L/min through

¹ This test method is under the jurisdiction of ASTM Committee D-22 on Sampling and Analysis of Atmospheres and is the direct responsibility of Subcommittee D22.03 on Ambient Atmospheres and Source Emissions.

Current edition approved May 10, 1998. Published July 1998. Originally published as D 6209-97. Last previous edition D 6209-97.

² This test method is based on U. S. Environmental Protection Agency Compendium Method TO-13, Compendium of Methods for the Determination of Toxic Organic Compounds in Ambient Air, Report No. EPA/600-4-89/018, June 1988, available from the National Technical Information Service, 5285 Port Royal Rd., Springfield, VA 22161, Order No. PB90-11989/AS.

³ *Annual Book of ASTM Standards*, Vol 11.03.

⁴ *Annual Book of ASTM Standards*, Vol 14.03.