



Designation: D6822 – 24



Manual of Petroleum Measurement Standards (MPMS), Chapter 9.3

Standard Test Method for Density, Relative Density, and API Gravity of Crude Petroleum and Liquid Petroleum Products by Thermohydrometer Method¹

This standard is issued under the fixed designation D6822; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope*

1.1 This test method covers the determination, using a glass thermohydrometer and/or a separate (alternative) temperature measuring device in conjunction with a series of calculations, of the density, relative density, or API gravity of crude petroleum, petroleum products, or mixtures of petroleum and nonpetroleum products normally handled as liquids and having Reid vapor pressures (Test Method **D323**) of 101.325 kPa (14.696 psia) or less. Values are determined at existing temperatures and corrected to 15 °C (or 60 °F) by means of a series of calculations and international standard tables. Alternative temperature measuring devices shall be either the appropriate liquid-in-glass ASTM thermometers meeting either Specifications **E1** or **E2251** or digital contact thermometers (Guides **E2877** and **D8164**).

1.2 The initial thermohydrometer readings obtained are uncorrected hydrometer readings and not density measurements. Readings are measured on a thermohydrometer at either the reference temperature or at another convenient temperature, and readings are corrected for the meniscus effect, the thermal glass expansion effect, alternate calibration temperature effects and to the reference temperature by means of calculations and Adjunct to **D1250** Guide for Use of the Petroleum Measurement Tables (API *MPMS* Chapter 11.1). Values measured (determined) at other than the reference temperature being hydrometer/thermohydrometer readings and not density measurements

1.3 Readings determined as density, relative density, or API gravity can be converted to equivalent values in the other units

or alternate reference temperatures by means of Interconversion Procedures (API *MPMS* Chapter 11.5) or Adjunct to **D1250** Guide for Use of the Petroleum Measurement Tables (API *MPMS* Chapter 11.1), or both, or tables as applicable.

1.4 The initial thermohydrometer reading (including temperature) shall be recorded before performing any calculations. The calculations required in **9.1.5** and Section **10** shall be applied to the initial thermohydrometer reading with observations and results reported as required by Section **11** prior to use in a subsequent calculation procedure (measurement ticket calculation, meter factor calculation, or base prover volume determination).

1.5 **Annex A1 – Annex A3** contain procedures for determining fit-for-service, verifying, or certifying the equipment for this test method.

1.6 The values stated in SI units are to be regarded as standard.

1.6.1 *Exception*—The values given in parentheses (inch-pound) are for information only.

1.7 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.8 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

¹ This test method is under the jurisdiction of ASTM Committee **D02** on Petroleum Products, Liquid Fuels, and Lubricants and the API Committee on Petroleum Measurement, and is the direct responsibility of Subcommittee **D02.02** /COMQ, the joint ASTM-API Committee on Hydrocarbon Measurement for Custody Transfer (Joint ASTM-API). This test method has been approved by the sponsoring committees and accepted by the Cooperating Societies in accordance with established procedures.

Current edition approved July 1, 2024. Published November 2024. Originally approved in 2002. Last previous edition approved in 2023 as D6822 – 23. DOI: 10.1520/D6822-24.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at www.astm.org/contact. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

*A Summary of Changes section appears at the end of this standard