



Standard Practice for Estimation of Measurement Uncertainty for Data from Regression-based Methods¹

This standard is issued under the fixed designation D7366; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice establishes a standard for computing the measurement uncertainty for applicable test methods in Committee D19 on Water. The practice does not provide a single-point estimate for the entire working range, but rather relates the uncertainty to concentration. The statistical technique of regression is employed during data analysis.

1.2 Applicable test methods are those whose results come from regression-based methods and whose data are intra-laboratory (not inter-laboratory data, such as result from round-robin studies). For each analysis conducted using such a method, it is assumed that a fixed, reproducible amount of sample is introduced.

1.3 Calculation of the measurement uncertainty involves the analysis of data collected to help characterize the analytical method over an appropriate concentration range. Example sources of data include: 1) calibration studies (which may or may not be conducted in pure solvent), 2) recovery studies (which typically are conducted in matrix and include all sample-preparation steps), and 3) collections of data obtained as part of the method's ongoing Quality Control program. Use of multiple instruments, multiple operators, or both, and field-sampling protocols may or may not be reflected in the data.

1.4 In any designed study whose data are to be used to calculate method uncertainty, the user should think carefully about what the study is trying to accomplish and much variation should be incorporated into the study. General guidance on designing studies (for example, calibration, recovery) is given in Appendix A. Detailed guidelines on sources of variation are outside the scope of this practice, but general points to consider are included in Appendix B, which is not intended to be exhaustive. With any study, the user must think carefully about the factors involved with conducting the

analysis, and must realize that the computed measurement uncertainty will reflect the quality of the input data.

1.5 Associated with the measurement uncertainty is a user-chosen level of statistical confidence.

1.6 At any concentration in the working range, the measurement uncertainty is plus-or-minus the half-width of the prediction interval associated with the regression line.

1.7 It is assumed that the user has access to a statistical software package for performing regression. A statistician should be consulted if assistance is needed in selecting such a program.

1.8 A statistician also should be consulted if data transformations are being considered.

1.9 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 *ASTM Standards*:²

D1129 [Terminology Relating to Water](#)

3. Terminology

3.1 *Definitions of Terms Specific to This Standard*:

3.1.1 *confidence level*—the probability that the prediction interval from a regression estimate will encompass the true value of the amount or concentration of the analyte in a subsequent measurement. Typical choices for the confidence level are 99 % and 95 %.

3.1.2 *fitting technique*—a method for estimating the parameters of a mathematical model. For example, ordinary least squares is a fitting technique that may be used to estimate the parameters $a_0, a_1, a_2, \dots$ of the polynomial model $y = a_0 + a_1x + a_2x^2 + \dots$, based on observed $\{x, y\}$ pairs. Weighted least squares is also a fitting technique.

¹ This practice is under the jurisdiction of ASTM Committee D19 on Water and is the direct responsibility of Subcommittee D19.02 on Quality Systems, Specification, and Statistics.

Current edition approved Jan. 1, 2013. Published January 2013. Originally approved in 2008. Last previous approval in 2008 as D7366 – 08. DOI: 10.1520/D7366-08R13.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.