



Designation: D7840 – 24

Standard Test Method for Foaming Tendencies of Non-Aqueous Engine Coolants in Glassware¹

This standard is issued under the fixed designation D7840; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers a simple glassware test for evaluating the tendency of non-aqueous engine coolants to foam under laboratory controlled conditions of aeration and temperature.

1.2 *Units*—The values stated in SI units are to be regarded as the standard. No other units of measurement are included in this standard.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.* For specific warning statements, see 7.2 and 7.3.

1.4 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

D1193 Specification for Reagent Water

D4725 Terminology for Engine Coolants and Related Fluids

E1 Specification for ASTM Liquid-in-Glass Thermometers

E128 Test Method for Maximum Pore Diameter and Permeability of Rigid Porous Filters for Laboratory Use

E230/E230M Specification for Temperature-Electromotive Force (emf) Tables for Standardized Thermocouples

¹ This test method is under the jurisdiction of ASTM Committee D15 on Engine Coolants and Related Fluids and is the direct responsibility of Subcommittee D15.06 on Glassware Performance Tests.

Current edition approved Sept. 1, 2024. Published October 2024. Originally approved in 2012. Last previous edition approved in 2017 as D7840 – 12 (2017). DOI: 10.1520/D7840-24.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at www.astm.org/contact. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

3. Terminology

3.1 *Definitions*—For definitions of general terms used in D15 standards, refer to Terminology D4725.

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *break time, n*—time required for the foam to collapse (after the air supply has been shut off) to the first appearance of an “eye” on the surface of the test solution.

3.2.2 *eye, n*—appearance of foam-free area on the surface of the test coolant surrounded by a ring of foam clinging to the cylinder walls.

4. Summary of Test Method

4.1 The non-aqueous coolant of interest is blown with air at a constant rate for 5 min while maintained at a constant temperature of $88\text{ }^{\circ}\text{C} \pm 1\text{ }^{\circ}\text{C}$ by means of a suitable temperature bath. The volume of foam and the time for such foam to break are measured.

5. Significance and Use

5.1 In the test method, coolants generally will be distinguished that have a tendency to foam excessively from those that are suitable for further evaluation to determine performance in actual service.

NOTE 1—In use, the foaming tendency of a coolant solution may be increased by service aging or contamination. A properly functioning pressure cap will tend to suppress foaming in coolant solutions.

6. Apparatus

6.1 *Container*, a 500 mL graduated container of heat-resistant glass having a diameter of 45 mm to 50 mm and a length of 380 mm.

6.2 *Temperature Bath*, a heat-resistant glass container large enough to permit immersion of the graduated container at least to the 350 mL graduation mark. A 4000 mL beaker is satisfactory.

6.3 *Heat Source*, any heating system capable of maintaining a uniform bath temperature of $\pm 1\text{ }^{\circ}\text{C}$. A 750 W electric hot plate is satisfactory.