



Designation: E83 – 16

Standard Practice for Verification and Classification of Extensometer Systems¹

This standard is issued under the fixed designation E83; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the U.S. Department of Defense.

1. Scope*

1.1 This practice covers procedures for the verification and classification of extensometer systems, but it is not intended to be a complete purchase specification. The practice is applicable only to instruments that indicate or record values that are proportional to changes in length corresponding to either tensile or compressive strain. Extensometer systems are classified on the basis of the magnitude of their errors.

1.2 Because strain is a dimensionless quantity, this document can be used for extensometers based on either SI or US customary units of displacement.

NOTE 1—Bonded resistance strain gauges directly bonded to a specimen cannot be calibrated or verified with the apparatus described in this practice for the verification of extensometers having definite gauge points. (See procedures as described in Test Methods E251.)

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:²

E6 Terminology Relating to Methods of Mechanical Testing
E21 Test Methods for Elevated Temperature Tension Tests of Metallic Materials

E251 Test Methods for Performance Characteristics of Metallic Bonded Resistance Strain Gages

2.2 Other Standards:³

JCGM 100:2008 Evaluation of measurement data – Guide to the expression of uncertainty in measurement

¹ This practice is under the jurisdiction of ASTM Committee E28 on Mechanical Testing and is the direct responsibility of Subcommittee E28.01 on Calibration of Mechanical Testing Machines and Apparatus.

Current edition approved Dec. 15, 2016. Published January 2017. Originally approved in 1950. Last previous edition approved in 2010 as E83 – 10a. DOI: 10.1520/E0083-16.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from International Organization for Standardization (ISO), ISO Central Secretariat, BIBC II, Chemin de Blandonnet 8, CP 401, 1214 Vernier, Geneva, Switzerland, <http://www.iso.org>.

3. Terminology

3.1 Definitions:

3.1.1 In addition to the terms listed, see Terminology E6.

3.1.2 *calibration*—a determination of the calibration factor for a system using established procedures.

3.1.3 *calibration factor*—the factor by which the change in extensometer reading must be multiplied to obtain the equivalent strain.

3.1.3.1 *Discussion*—For any extensometer, the calibration factor is equal to the ratio of change in length to the product of the gauge length and the change in the extensometer reading. For direct-reading extensometers the calibration factor is unity.

3.1.4 *compressometer*—a specialized extensometer used for sensing negative or compressive strain.

3.1.5 *deflectometer*—a specialized extensometer used for sensing of extension or motion, usually without reference to a specific gauge length.

3.1.6 *error, in extensometer systems*—the value obtained by subtracting the correct value of the strain from the indicated value given by the extensometer system.

3.1.7 *extensometer, n*—a device for sensing strain.

3.1.8 *extensometer systems*—a system for sensing and indicating strain.

3.1.8.1 *Discussion*—The system will normally include an extensometer, conditioning electronics and auxiliary device (recorder, digital readout, computer, etc.). However, completely self-contained mechanical devices are permitted. An extensometer system may be one of three types.

3.1.9 *Type 1 extensometer system, n*—an extensometer system which both defines gauge length and senses extension, for example, a clip-on strain gauge type with conditioning electronics.

3.1.10 *Type 2 extensometer system, n*—an extensometer which senses extension and the gauge length is defined by specimen geometry or specimen features such as ridges or notches.

3.1.10.1 *Discussion*—A Type 2 extensometer is used where the extensometer gauge length is determined by features on the specimen, for example, ridges, notches, or overall height (in case of compression test piece). The precision associated with

*A Summary of Changes section appears at the end of this standard