



Designation: E 1045 – ~~85~~ (Reapproved 1996) 00

Standard Specification for Pipet, Sahli Hemoglobin¹

This standard is issued under the fixed designation E 1045; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This specification covers reusable pipets calibrated“ to contain” 20 cmm of whole blood and used for hemoglobin determinations.

2. Materials

~~2.1 The pipets shall be made~~Referenced Documents

2.1 ASTM Standards:

¹ This specification is under the jurisdiction of ASTM Committee E41 on Laboratory Apparatus and is the direct responsibility of Subcommittee E41.071 on Microchemical Apparatus.

Current edition approved ~~Jan. 25, 1985~~; Nov. 10, 2000. Published ~~March 1985~~; January 2001. Originally published as E 1045 – 85. Last previous edition E 1045 – 85.