



Standard Guide for Describing and Specifying the Spectrometer of an Optical Emission Direct-Reading Instrument¹

This standard is issued under the fixed designation E 1507; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This guide covers features of a spectrometer or polychromator used for optical emission, direct-reading, spectrochemical analysis. A polychromator in this sense consists of a spectrometer with an extended and fixed wavelength range and an array of fixed exit slits to isolate the spectral lines of the elements to be measured.

1.1.1 This guide does not apply to direct-reading systems that employ echelle spectrometers and vidicon or other detectors where the design parameters are quite different.

1.2 This guide covers only the optical portion of the instrument, from excitation stand to photomultipliers.

1.2.1 Only general statements are made about source units.

1.2.2 Photomultipliers are included to the extent that they are mounted within the spectrometer to convert optical intensities to electrical signals, and establish the instrumental precision of each channel as a light measuring device. Readout systems are not included.

1.3 It is not the purpose of this guide to establish binding specifications or tolerances, but rather, to call attention to important parameters that manufacturers should include in their literature, to provide methods for measuring those parameters, and to assign values that are indicative of acceptably good performance. Because of the great variety of demands imposed by spectrochemical techniques, rigid performance criteria are not feasible.

1.4 The values stated in SI units are to be regarded as the standard.

1.5 *This standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

1.6 A partial listing of the information in this guide includes the following:

Section

Terminology	3
Background Equivalent Concentration (BEC)	3.2.1
Channel	3.2.2
Readability	3.2.3
Scattered Radiation	3.2.4
Shot Noise	3.2.5
White Light Precision	3.2.6
Working Resolution	3.2.7
Fundamental Spectrochemical Objectives	5
Accuracy	5.2
Precise Photometry	5.2.1
Bias in Photometry	5.2.2
Detection and Determination Limits	5.3
Minimization of Interference	5.4
Parameters of Spectrometer and Associated Components	6
Dispersion, Reciprocal Linear	6.2
Working Resolution	6.3
Speed or Radiation Throughput	6.4
White Light Precision	6.5
Scattered Light	6.6
Slits	6.7
Accuracy of Positioning Exit Slits	6.7.3
Wavelength Coverage and Focal Curve Length	6.8
Secondary Optics	6.9
Optical Stability	6.10
Spectrometer Illumination	6.11
Astigmatic Image	6.12
Excitation Stands	6.13
Vacuum Systems	6.14
Flushing With Transparent Gas (Nitrogen)	6.15
Other, Including Maintenance Features	6.16-6.18
Measuring Specified Polychromator Parameters	7
Working Resolution	7.1-7.4
Line Interference	7.5
White Light Precision	7.6
Scattered Light	7.7
Optical Alignment and Focus	7.8
Optical Stability	7.9
Precision and Accuracy	7.10
Describing the Spectrometer in Analytical Methods	8
Wavelength Coverage and Reciprocal Dispersion	8.2.1
Working Resolution	8.2.2
Entrance and Exit Slit Widths	8.2.3
Low Concentrations	8.2.4
High Concentrations	8.2.5
Varying Parameters on Working Resolution and Throughput	Appendix X1
Dispersion	X1.2
Size of Spectrometer	X1.3
Effects of Slit Widths	X1.4
Recommended Spectral Lines for Measuring Working Resolution	Appendix X2
Photographic Speed Versus Photoelectric Throughput	Appendix X3
Photographic Speed	X3.1
Polychromator Radiation Throughput	X3.2
Calculation of Radiation Throughput	X3.3

¹ This guide is under the jurisdiction of ASTM Committee E01 on Analytical Chemistry for Metals, Ores, and Related Materials and is the direct responsibility of Subcommittee E01.20 on Fundamental Practices.

Current edition approved June 10, 2003. Published September 2003. Originally approved in 1993. Last previous edition approved in 1998 as E 1507 - 98.