



Standard Test Method for Ignitable Liquid Residues in Extracts from Fire Debris Samples by Gas Chromatography-Mass Spectrometry¹

This standard is issued under the fixed designation E1618; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the identification of residues of ignitable liquids in extracts from fire debris samples. Extraction procedures are described in the referenced documents.

1.2 Although this test method is suitable for all samples, it is especially appropriate for extracts that contain high background levels of substrate materials or pyrolysis and combustion products. This test method is also suitable for the identification of single compounds, simple mixtures, or non-petroleum based ignitable liquids.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 This practice cannot replace knowledge, skill, or ability acquired through appropriate education, training, and experience and should be used in conjunction with sound professional judgment.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 *ASTM Standards:*²

[E1386 Practice for Separation of Ignitable Liquid Residues from Fire Debris Samples by Solvent Extraction](#)

[E1388 Practice for Sampling of Headspace Vapors from Fire Debris Samples](#)

[E1412 Practice for Separation of Ignitable Liquid Residues from Fire Debris Samples by Passive Headspace Concentration With Activated Charcoal](#)

[E1413 Practice for Separation of Ignitable Liquid Residues from Fire Debris Samples by Dynamic Headspace Concentration](#)

[E2154 Practice for Separation and Concentration of Ignitable Liquid Residues from Fire Debris Samples by Passive Headspace Concentration with Solid Phase Microextraction \(SPME\)](#)

[E2451 Practice for Preserving Ignitable Liquids and Ignitable Liquid Residue Extracts from Fire Debris Samples](#)

3. Summary of Test Method

3.1 The sample is analyzed with a gas chromatograph (GC) which is interfaced to a mass spectrometer (MS) and a data system (DS) capable of storing and manipulating chromatographic and mass spectral data.

3.2 Post-run data analysis generates extracted ion profiles (mass chromatograms) characteristic of the chemical compound types commonly found in ignitable liquids. Additionally, specific chemical components (target compounds) may be identified by their mass spectra and retention times. Semi-quantitative determination of target compounds which are identified by mass spectra and retention time may be used to develop target compound chromatograms (TCCs).

3.2.1 The total ion chromatogram (TIC), extracted ion profiles (EIP) for the alkane, alkene, alcohol, aromatic, cycloalkane, ester, ketone and polynuclear aromatic compound types, or TCCs, or combination thereof, are evaluated by visual pattern matching against known reference ignitable liquids.

3.2.2 Ignitable liquids may be grouped into one of seven major classifications or one miscellaneous class, as described in this test method.

4. Significance and Use

4.1 The identification of an ignitable liquid residue in samples from a fire scene can support the field investigator's opinion regarding the origin, fuel load, and incendiary nature of the fire.

4.1.1 The identification of an ignitable liquid residue in a fire scene does not necessarily lead to the conclusion that a fire was incendiary in nature. Further investigation may reveal a legitimate reason for the presence of ignitable liquid residues.

¹ This test method is under the jurisdiction of ASTM Committee E30 on Forensic Sciences and is the direct responsibility of Subcommittee E30.01 on Criminalistics.

Current edition approved July 1, 2014. Published July 2014. Originally approved in 1994. Last previous edition approved in 2011 as E1618 – 11. DOI: 10.1520/E1618-14.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.