



Standard Specification for Miniature Thermocouple Connectors¹

This standard is issued under the fixed designation E1684/E1684M; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This specification covers separable single-circuit miniature thermocouple connectors with two flat pins. Connectors covered by this specification must be rated for continuous use to at least 300°F (150°C), but they may optionally be rated to a higher temperature.

1.2 This specification does not cover multiple-circuit connectors, multi-pin connectors, standard size connectors, or connectors intended primarily for panel mounting. High-temperature connectors (for example, those designed for continuous use at temperatures significantly above approximately 500°F [260°C]) are not intended to be covered by this specification.

1.3 The values stated in either SI units or inch-pound units are to be regarded separately as standard. The values stated in brackets are not conversions to the values they succeed and therefore shall be used independently of the preceding values. The values given in parentheses are conversions of the values they succeed. Combining values from the two systems may result in non-conformance with the standard.

1.4 The plug and jack may be ordered separately but many of the connector characteristics (see [Table 1](#)) are qualified as a mated pair. The requirements in this standard are applicable for the connector pair supplied by the same manufacturer.

1.5 The following precautionary statement pertains only to the Test Methods portion, Section 9, of this specification: *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 The following documents of the latest issue form a part of this specification to the extent referenced herein. In case of

conflict between this specification and another referenced document, this specification shall take precedence.

2.2 ASTM Standards:²

[E230/E230M Specification and Temperature-Electromotive Force \(emf\) Tables for Standardized Thermocouples](#)

[E344 Terminology Relating to Thermometry and Hydrometry](#)

[E2820 Test Method for Evaluating Thermal EMF Properties of Base-Metal Thermocouple Connectors](#)

3. Terminology

3.1 *Definitions*—The definitions given in Terminology [E344](#) shall apply.

3.2 Definitions of Terms Specific to This Standard:

3.2.1 *connector pair, n*—assembly consisting of a plug and a jack, each having both positive and negative inserts, that will connect two parts of an electrical circuit and provide a means of physically disconnecting the two parts without the use of a tool.

3.2.2 *contact insert, n*—metallic conductor assemblies that, when installed in connector bodies, provide connections between two parts of an electrical circuit. Plug connectors will contain projecting prong contacts, while jack connectors will contain recessed socket or receptacle contacts.

3.2.3 *service life, n*—period of time over which a connector assembly can be put to use and retain all physical and thermoelectric properties.

3.2.4 *test difference, n*—thermoelectric emf value attributable to mated connectors observed by the test procedure of this specification.

4. Significance and Use

4.1 The widespread use of thermocouple connectors requires standardization of mating dimensions and performance characteristics.

4.2 This specification describes miniature thermocouple connector dimensions and capabilities and includes test procedures suitable for evaluating the performance of a particular

¹ This specification is under the jurisdiction of ASTM Committee E20 on Temperature Measurement and is the direct responsibility of Subcommittee E20.04 on Thermocouples.

Current edition approved Dec. 1, 2015. Published January 2016. Originally approved in 1995. Last previous edition approved in 2010 as E1684 – 10. DOI: 10.1520/E1684-15.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.