



Standard Guide for the Preparation and Evaluation of Liquid Baths Used for Temperature Calibration by Comparison¹

This standard is issued under the fixed designation E2488; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

INTRODUCTION

Many of the Standards and Test Methods under the jurisdiction of ASTM committee E20 on Temperature Measurement make reference to the use of controlled temperature fluid baths for the calibration of thermometers by the comparison method. In this method the thermometer under test is measured while immersed in an isothermal medium whose temperature is simultaneously determined by a calibrated reference thermometer. The uncertainty of all such comparison calibrations depends upon how well the isothermal conditions can be maintained. The bath temperature must be stable over time and uniform within the working space at the operating temperatures. This guide provides basic information, options and instructions that will enable the user to prepare and evaluate controlled temperature baths for calibrations.

1. Scope

1.1 This guide is intended for use with controlled temperature comparison baths that contain test fluids and operate within the temperature range of -100°C to 550°C .

1.2 This guide describes the essential features of controlled temperature fluid baths used for the purpose of thermometer calibration by the comparison method.

1.3 This guide does not address the details on the design and construction of controlled-temperature fluid baths.

1.4 This guide describes a method to define the working space of a bath and evaluate the temperature variations within this space. Ideally, the working space will be as close as possible to isothermal.

1.5 This guide does not address fixed point baths, ice point baths or vapor baths.

1.6 This guide does not address fluidized powder baths.

1.7 This guide does not address baths that are programmed to change temperature.

1.8 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.9 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:²

- E1 Specification for ASTM Liquid-in-Glass Thermometers
- E344 Terminology Relating to Thermometry and Hydrometry
- E644 Test Methods for Testing Industrial Resistance Thermometers
- E839 Test Methods for Sheathed Thermocouples and Sheathed Thermocouple Cable

2.2 Other Documents:

- ITS-90 The International Temperature Scale of 1990³
- NIST Monograph 126 Platinum Resistance Thermometry⁴
- NIST Monograph 150 Liquid-in-Glass Thermometry⁴
- NIST SP 250-22 Platinum Resistance Thermometer Calibrations⁴

¹ This guide is under the jurisdiction of ASTM Committee E20 on Temperature Measurement and is the direct responsibility of Subcommittee E20.07 on Fundamentals in Thermometry.

Current edition approved Dec. 1, 2014. Published December 2014. Originally approved in 2009. Last previous edition approved in 2009 as E2488 – 09. DOI: 10.1520/E2488-09R14.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Preston-Thomas, H., *METROLOGIA*, Vol. 27, 1990, pp 3-10 and 107 (errata). Mangum, B. W., *JOURNAL OF RESEARCH*, National Institute of Standards and Technology, Vol 95, 1990, p. 69.

⁴ Available from National Institute of Standards and Technology (NIST), 100 Bureau Dr., Stop 1070, Gaithersburg, MD 20899-1070, <http://www.nist.gov>.