



Designation: F 1472 – 02a

Standard Specification for Wrought Titanium-6Aluminum-4Vanadium Alloy for Surgical Implant Applications (UNS R56400)¹

This standard is issued under the fixed designation F 1472; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope*

1.1 This specification covers the chemical, mechanical, and metallurgical requirements for wrought annealed titanium-6aluminum-4vanadium alloy (UNS R56400) to be used in the manufacture of surgical implants.

1.2 The values stated in inch-pound units are to be regarded as the standard. The SI equivalents in parentheses are provided for information only.

2. Referenced Documents

2.1 *ASTM Standards:*

E 8 Test Methods of Tension Testing of Metallic Materials²

¹ This specification is under the jurisdiction of ASTM Committee F04 on Medical and Surgical Materials and Devices and is the direct responsibility of Subcommittee F04.12 on Metallurgical Materials.

Current edition approved Apr. Nov. 10, 2002. Published June 2002. January 2003. Originally published as F 1472—93, in 1993. Last previous edition approved in 2002 as F 1472 – 002.

*A Summary of Changes section appears at the end of this standard.