



Standard Guide for Design Verification Device Size and Sample Size Selection for Endovascular Devices¹

This standard is issued under the fixed designation F3172; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This guide provides guidance for selecting an appropriate device size(s) and determining an appropriate sample size(s) (that is, number of samples) for design verification testing of endovascular devices. A methodology is presented to determine which device size(s) should be selected for testing to verify the device design adequately for each design input requirement (that is, test characteristic). Additionally, different statistical approaches are presented and discussed to help guide the developer to determine and justify sample size(s) for the design input requirement being verified. Alternate methodologies for determining device size selection and sample size selection may be acceptable for design verification.

1.2 This guide applies to physical design verification testing. This guide addresses in-vitro testing; in-vivo/animal studies are outside the scope of this guide. This guide does not directly address design validation; however, the methodologies presented may be applicable to in-vitro design validation testing. Guidance for sampling related to computational simulation (for example, sensitivity analysis and tolerance analysis) is not provided. Guidance for using models, such as design of experiments (DOE), for design verification testing is not provided. This guide does not address sampling across multiple manufacturing lots as this is typically done as process validation. Special considerations are to be given to certain tests such as fatigue (see Practice E739) and shelf life testing (see Section 8).

1.3 Regulatory guidance may exist for endovascular devices that should be considered for design verification device size and sample size selection.

1.4 *Units*—The values stated in SI units are to be regarded as the standard. No other units of measurement are included in this standard.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the*

responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

2. Referenced Documents

2.1 ASTM Standards:²

E739 Practice for Statistical Analysis of Linear or Linearized Stress-Life ($S-N$) and Strain-Life ($\epsilon-N$) Fatigue Data

F2914 Guide for Identification of Shelf-life Test Attributes for Endovascular Devices

2.2 ISO Standards:³

ISO 14971:2012 Medical devices—Application of risk management to medical devices

3. Terminology

3.1 Definitions:

3.1.1 *attribute data, n*—data that identify the presence or absence of a characteristic (for example, good/bad or pass/fail).

3.1.2 *design input requirements, n*—physical and performance requirements of a device that are used as a basis for device design (typically defined as test characteristics such as balloon burst pressure, shaft tensile strength, and so forth).

3.1.3 *design output, n*—features of the device (that is, dimensions, materials, and so forth) that define the design and make it capable of achieving design input requirements.

3.1.4 *design subgroup, n*—set defined by the device sizes within the device matrix in which the essential design outputs do not vary for a specified design input requirement (that is, device sizes that share the same design for a specified design input requirement).

3.1.5 *design validation, n*—establishing by objective evidence that the device conforms to defined user needs and intended use(s).

¹ This guide is under the jurisdiction of ASTM Committee F04 on Medical and Surgical Materials and Devices and is the direct responsibility of Subcommittee F04.30 on Cardiovascular Standards.

Current edition approved Dec. 15, 2015. Published February 2016. DOI: 10.1520/F3172-15.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from American National Standards Institute (ANSI), 25 W. 43rd St., 4th Floor, New York, NY 10036, <http://www.ansi.org>.