



Standard Test Method for Calibration of Narrow- and Broad-Band Ultraviolet Radiometers Using a Spectroradiometer¹

This standard is issued under the fixed designation G130; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

INTRODUCTION

Accurate and precise measurements of ultraviolet irradiance are required in the determination of the radiant exposure of both total and selected narrow bands of ultraviolet radiation for the determination of exposure levels in (1) outdoor weathering of materials, (2) indoor accelerated exposure testing of materials using manufactured light sources, and (3) UV-A and UV-B ultraviolet radiation in terms both of the assessment of climatic parameters and the changes that may be taking place in the solar ultraviolet radiation reaching earth.

Although meteorological measurements usually require calibration of pyranometers and radiometers oriented with axis vertical, applications associated with materials testing require an assessment of the calibration accuracy at orientations with the axis horizontal (usually associated with testing in indoor exposure cabinets) or with the axis at angles typically up to 45° or greater from the horizontal (for outdoor exposure testing). These calibrations also require that deviations from the cosine law, tilt effects, and temperature sensitivity be either known and documented for the instrument model or determined on individual instruments.

This test method requires calibrations traceable to primary reference standards maintained by a national metrological laboratory that has participated in intercomparisons of standards of spectral irradiance.

1. Scope

1.1 This test method covers the calibration of ultraviolet light-measuring radiometers possessing either narrow- or broad-band spectral response distributions using either a scanning or a linear-diode-array spectroradiometer as the primary reference instrument. For transfer of calibration from radiometers calibrated by this test method to other instruments, Test Method E824 should be used.

NOTE 1—Special precautions must be taken when a diode-array spectroradiometer is employed in the calibration of filter radiometers having spectral response distributions below 320-nm wavelength. Such precautions are described in detail in subsequent sections of this test method.

1.2 This test method is limited to calibrations of radiometers against light sources that the radiometers will be used to measure during field use.

NOTE 2—For example, an ultraviolet radiometer calibrated against natural sunlight cannot be employed to measure the total ultraviolet irradiance of a fluorescent ultraviolet lamp.

1.3 Calibrations performed using this test method may be against natural sunlight, Xenon-arc burners, metal halide burners, tungsten and tungsten-halogen lamps, fluorescent lamps, etc.

1.4 Radiometers that may be calibrated by this test method include narrow-, broad-, and wide-band ultraviolet radiometers, and narrow-, broad, and wide-band visible-region-only radiometers, or radiometers having wavelength response distributions that fall into both the ultraviolet and visible regions.

NOTE 3—For purposes of this test method, narrow-band radiometers are those with $\Delta\lambda \leq 20$ nm, broad-band radiometers are those with $20 \text{ nm} \leq \Delta\lambda \leq 70$ nm, and wide-band radiometers are those with $\Delta\lambda \geq 70$ nm.

NOTE 4—For purposes of this test method, the ultraviolet region is defined as the region from 285 to 400-nm wavelength, and the visible region is defined as the region from 400 to 750-nm wavelength. The ultraviolet region is further defined as being either UV-A with radiation of wavelengths from 315 to 400 nm, or UV-B with radiation from 285 to 315-nm wavelength.

¹ This test method is under the jurisdiction of ASTM Committee G03 on Weathering and Durability and is the direct responsibility of Subcommittee G03.09 on Radiometry.

Current edition approved June 1, 2012. Published January 2013. Originally approved in 1995. Last previous edition approved in 2006 as G130–06 DOI: 10.1520/G0130-12.